Int. j. adv. multidisc. res. stud. 2025; 5(4):1624-1631

Received: 03-07-2025
Accepted: 13-08-2025

International Journal of Advanced Multidisciplinary Research and Studies

ISSN: 2583-049X

Operational Mechanics and Engineering Concepts in Diagnostic X-ray and CT Imaging Systems

[1] Aunde Philip Fanen, [2] Tseaa Sekaa Thomas, [3] Victor Uchenna Uchendu, [4] Akpotu Efe, [5] Nwigbo Hyginus Peter, [6] Nduul Johnson Terdue, [7] Achu Paschal Aondohemba, [8] Nsikan Kufre Akpan, [9] Ugwuanyi Desmond Chinaemere, [10] Odo Bonaventure Ewezuga

[1] Department of Physics, Joseph Sarwuan Tarka University, Makurdi, Benue State, Nigeria
[2, 6] Department of Industrial Physics, Joseph Sarwuan Tarka University, Makurdi, Benue State, Nigeria
[3] Department of Civil Engineering, Kwara State Polytechnic, Ilorin, Kwara State, Nigeria
[4] Oil and Gas Engineering, Academy of Engineering, RUDN University, Moscow, Russia
[5] Glorious Saint International College, Ikpata Noyo, Alike Ikwo, Ebonyi State, Nigeria
[7] Department of Chemistry Joseph Sarwuan Tarka University Makurdi Benue State Nigeria (former University of Agriculture Makurdi Benue state Nigeria)
[8] Department of Surveying and Geoinformatics, Gregory University Uturu, Abia State, Nigeria
[9] Computer Science, Osisatech Secondary School Enugu, Enugu State, Nigeria
[10] Computer Science, Community Secondary School Ibenda, Enugu State, Nigeria

Corresponding Author: **Aunde Philip Fanen**

Abstract

Diagnostic X-ray and Computed Tomography (CT) imaging systems are fundamental tools in modern medical diagnostics, relying on the integration of operational mechanics and engineering principles to deliver high-quality clinical images. The core mechanics of X-ray generation involve the acceleration of electrons within an X-ray tube and their interaction with a metal target, producing bremsstrahlung and characteristic radiation. Engineering concepts such as tube voltage, current regulation, and filtration directly influence image quality, patient dose, and diagnostic accuracy. In CT systems, the principles of rotational mechanics, slip-ring technology, and detector array design enable continuous data acquisition, facilitating rapid volumetric imaging. Advanced engineering solutions, including beam collimation, anti-scatter grids, and dual-energy imaging, further optimize contrast resolution while minimizing radiation exposure. Signal detection and digital image reconstruction employ mathematical algorithms, such as filtered back projection and iterative reconstruction, to enhance spatial resolution and noise suppression. Moreover, mechanical precision in gantry rotation, detector alignment, and patient positioning plays a pivotal role in ensuring consistency and reliability of diagnostic output. Integration of dose reduction strategies, such as automatic exposure control and adaptive filtration, reflects the balance between engineering innovation and radiation safety standards. Understanding the interplay of operational mechanics with applied engineering not only informs system design but also supports advancements in image quality, efficiency, and patient care. This synthesis of physics and engineering underscores the evolving capabilities of X-ray and CT systems in clinical diagnostics.

Keywords: Operational Mechanics, Engineering Concepts, Diagnostic X-ray, CT Imaging Systems

Introduction

Research by Vassileva and Rehani [1] stressed the necessity of monitoring patient radiation doses to avoid unnecessary exposure in diagnostic imaging. They stated that improved imaging technologies, notably CT scans, contribute to higher effective doses than traditional X-rays.

Ajayi and Akinwumiju [2] did a study in Nigeria and discovered significant differences in radiation doses between facilities due to uneven adherence to standard protocols. Their findings suggested that there is an urgent need for imaging practice training and harmonization in order to eliminate dosage discrepancies and increase patient safety.

Suleiman *et al.* [3] conducted study in Sub-Saharan Africa and found a lack of reliable data on organ doses and effective doses from diagnostic imaging. To provide radiation safety, the study recommended that dose reference levels (DRLs) be established that are customized to local populations. This is consistent with international recommendations from the International Commission on Radiological Protection, which promotes the use of DRLs as a dose management tool in medical imaging.

Alzimami *et al.* [4] studied how imaging parameters such as tube voltage and current affect patient dosages in CT imaging. Their findings showed that altering these factors could enable dosage optimization without compromising image quality, emphasizing the importance of equipment settings in radiation safety.

A study conducted by Nkubli *et al.* [5] in Ghana focused on the variability of effective doses among patients undergoing X-ray examinations. The authors reported significant disparities across facilities, which were attributed to differences in operator expertise and equipment calibration. They emphasized the need for regular training and the standardization of imaging practices to reduce dose variability.

Aliyu *et al.* [6] examined organ equivalent dosages in paediatric patients getting CT scans in Nigeria. The study found that youngsters were more vulnerable to radiation-induced impacts due to their heightened sensitivity. The authors suggested using paediatric-specific imaging procedures to reduce hazards while maintaining diagnostic accuracy.

Olowookere *et al.* [7] examined the effectiveness of automated dose monitoring systems in improving radiation safety in diagnostic imaging. Their results showed that these systems could provide real-time feedback, enabling radiologists to adjust imaging parameters dynamically to reduce exposure.

Diagnostic X-rays are widely used in medical imaging because they can provide critical information for patient care. However, these procedures expose patients to ionizing radiation, which can be harmful to their health. Accurate calculation of organ equivalent and effective dosages is critical for risk management, procedure optimization, and patient safety. Recent advances have focused on increasing the precision of dosage estimates in order to reduce adverse effects while maintaining diagnostic quality [8].

Recent research has used a variety of methods to estimate dosages, including direct measurement, mathematical modelling, and computer simulations. Direct measurement includes placing dosimeters on patients, whereas simulations employ computational models to estimate dose distribution within the body using X-ray characteristics [9].

One typical way for calculating doses is to use mathematical models that account for X-ray output, patient location, and specific machine settings like tube current (mAs) and voltage (kV). Monte Carlo simulation models have been very useful in forecasting dose distributions because they can mimic complicated interactions between X-rays and human tissues [10].

The development of automated software and its connection with Digital Imaging and Communications in Medicine (DICOM) data has made dose estimation easier. These devices provide real-time dose tracking and greater precision in organ dosage estimations. Current research focuses on standardizing these tools to ensure consistency across multiple healthcare facilities [11].

Monte Carlo methods have been frequently utilized to model radiation transport and interactions in the human body. These models account for complicated geometries and tissue compositions, resulting in highly precise dose estimations. Recent research has demonstrated that Monte Carlo-based algorithms can accurately estimate doses for a variety of diagnostic procedures, including computed tomography (CT) and fluoroscopy [12].

Several clinical trials have been carried out to gather information on organ and effective doses from regular diagnostic techniques. For example, research in Nigeria evaluated dosages for patients receiving routine X-ray exams like chest and abdominal imaging. The results showed the diversity in dosages due to changes in equipment settings and patient anatomy, underlining the importance of consistent protocols [13]. The ICRP has issued guidelines to limit the effective dose to patients, advocating the optimization of processes to reduce exposure. Their papers also emphasize the value of employing dose reference levels (DRLs) to benchmark and control radiation doses during diagnostic operations [14].

Children are more sensitive to radiation than adults, making accurate dose estimation even more crucial in paediatric imaging. Recent studies emphasize the need for tailored protocols that adjust parameters such as tube voltage and current to reduce the effective dose in paediatric patients without compromising image quality [15].

According to research, radiation doses vary significantly between countries due to differences in equipment, processes, and healthcare infrastructure. Studies comparing industrialized and developing nations indicate that effective dosage management is less stringent in the latter, resulting in greater average doses for equivalent operations [3].

Optimizing radiological procedures can drastically minimize the doses administered to patients. Strategies include increasing tube voltages, modifying patient posture, and implementing dose-saving technology such as automatic exposure control. These modifications can reduce the entrance surface dosage and the effective dose [16].

Automated dosage monitoring systems have been used in various healthcare settings to track and analyze patient doses in real time. These devices employ DICOM data to compute organ and effective doses, providing operators with quick feedback and helping to optimize protocols [17].

Despite advancements, it is still difficult to correctly estimate organ and effective doses. Patient movement, differences in body composition, and the intricacy of multi-phase imaging studies all have an impact on dosage estimations. Ongoing research focuses on improving models and software algorithms that solve these difficulties [18].

Effective Dose

Effective dose is a radiological statistic that estimates the overall risk of ionizing radiation exposure by taking into account both the kind of radiation and the sensitivity of each exposed organ. Unlike organ equivalent dose, which measures radiation received by a single organ, effective dose integrates doses across organs, taking into account their various sensitivities, to offer a single value that represents the risk for the entire body. Mettler *et al.* [11] developed effective dose to help radiologists analyze radiation risks from various diagnostic and therapeutic treatments, guiding them in improving protocols and minimizing patient exposure.

Effective dose is calculated using tissue weighting factors, which are assigned to organs based on their radiosensitivity. For instance, organs such as the lungs and bone marrow have higher weighting factors because of their increased susceptibility to radiation-induced cancer. This calculation allows healthcare professionals to estimate the cancer risk and other long-term effects associated with radiation exposure, making it a crucial measure in procedures like computed tomography (CT), X-rays, and nuclear medicine [19].

In medical imaging, effective dose is a universal statistic that may be used to compare doses across different imaging modalities and to standardize dosage limits across populations and age groups. Recent studies have shown that the effective dose varies greatly depending on the imaging technique; for example, CT scans often result in higher effective doses than normal radiography because CT scans frequently entail repeated exposures and cover a larger area of the body [20]. This parameter is critical for balancing the diagnostic benefits of imaging treatments with the possible dangers of radiation.

Pediatric imaging frequently uses modified procedures to lower effective dosage since children are more radiosensitive and have a longer risk window for acquiring radiation-induced disorders. According to Alameen *et al.* [21], dose modifications, such as lowering X-ray tube voltage and current, can dramatically reduce the effective dose in children while maintaining image quality. This method is becoming an increasingly important component of radiation safety guidelines for young patients.

The effective dose remains a cornerstone of radiological safety, enabling for better decision-making by comparing radiation risk across different operations. As technology advances, the accuracy and application of effective dosage measurements improve, supporting safer medical imaging procedures.

Radiation Units and Measurement Standards

Radiation units and measurement standards are vital for quantifying and assessing radiation exposure, especially in medical and nuclear applications. These units such as gray (Gy), sievert (Sv), and becquerel (Bq) enable professionals to evaluate absorbed dose, potential biological impact, and radioactivity, respectively. The establishment of these units by organizations like the International Commission on Radiological Protection (ICRP) and the International Atomic Energy Agency (IAEA) has standardized the practice of radiation safety worldwide, allowing for consistent assessment, monitoring, and regulation [11].

The gray (Gy) is a measurement of absorbed dose that measures the amount of radiation energy deposited in a medium, often human tissue, per unit mass. It is essential in radiotherapy, where precise dosages are required to enhance tumor control while avoiding damage to adjacent healthy tissues [22]. In clinical settings, measuring absorbed dosage is critical for predicting treatment outcomes and reducing hazards, especially considering the specific susceptibility of different tissues to radiation exposure [23].

Sievert (Sv) assesses equivalent and effective doses and is critical for evaluating the health hazards associated with radiation exposure. Equivalent dose is a radiation weighting factor that modifies the absorbed dosage to account for the various effects of different radiation types (e.g., alpha, beta, gamma). Effective dosage also takes into account tissue weighting parameters, resulting in a cumulative risk assessment across various organs based on radiosensitivity [24]. This technique is critical for radiologists and radiologic technologists, who employ effective dose values to assess and limit potential dangers from diagnostic imaging procedures like CT scans and fluoroscopy, which frequently involve higher radiation doses [25].

The becquerel (Bq) quantifies radioactivity, especially the rate at which a radioactive material undergoes nuclear decay, with one Bq equalling one disintegration per second. This unit is especially important in nuclear medicine, since radioactive decay rates influence both diagnostic efficacy and patient safety. In environmental monitoring, Bq is also used to measure radioactive contamination in areas affected by nuclear events, as demonstrated by research on radioactive fallout and waste management techniques [26].

Calibration and regulatory criteria are necessary to ensure accurate measurement and application of these units. The IAEA and WHO, among others, have produced standards for calibrating dosimetry equipment, allowing for consistent and trustworthy readings across facilities and applications. These standards encourage the safe use of radiation in medical and industrial settings, assisting specialists in adhering to dosage limitations that safeguard both patients and radiation workers. The International Commission on Radiological Protection (ICRP), for example, recommends annual dose limits of 20 mSv for radiation workers and 1 mSv for the general public to avoid potential health risks from cumulative exposure.

In recent years, advances in digital dosimeters and automated dose-tracking systems have improved radiation monitoring by delivering real-time dosage information, allowing healthcare personnel to change imaging techniques on the fly. These methods are extremely useful for managing cumulative doses, especially for patients having many imaging sessions or workers in high-radiation situations [27]. Furthermore, artificial intelligence (AI) is rapidly being used to improve dose management, with AI-powered software capable of recommending tailored radiation settings based on patient-specific data [28].

Techniques and Methods for Dose Estimation

Dose estimation techniques in radiology are crucial for calculating how much radiation patients absorb during diagnostic and therapeutic operations. These strategies aid in the optimization of radiation protection, risk management for patients, and treatment planning accuracy. Direct measuring methods, computer modelling, Monte Carlo simulations, and patient-specific modelling are examples of important approaches and methodologies.

1. Direct Measurement

Direct measuring techniques use dosimeters to assess the absorbed dosage in real time or after exposure. Dosimeters, such as thermos luminescent dosimeters (TLDs) and optically stimulated luminescence dosimeters (OSLDs), can be applied to or near the patient during exposure to measure organ dose. These technologies deliver accurate real-time data, but they require careful positioning and calibration to maintain precision [18].

2. Mathematical and Computational Model

Mathematical models use standardized anthropomorphic phantoms (models of the human body or specific organs) to compute radiation doses. Phantoms aid in dose calculation by modelling patient anatomy and radiation interaction with tissues, resulting in a more generalized dose estimate for typical patients. Although extensively used, these models may not adequately account for changes in individual patient anatomy [29].

3. Monte Carlo Stimulations

Monte Carlo simulations are among the most complex and effective methods of dose prediction, as they use probabilistic calculations to describe radiation particle interactions with biological tissues. Monte Carlo algorithms simulate millions of possible interactions, resulting in highly detailed dosage distributions for particular organs and tissues. These simulations are frequently used to estimate doses in complex radiographic procedures, such as CT and interventional radiology, where patient-specific characteristics and equipment settings influence dose distribution [30].

4. Patient Specific Modelling

Patient-specific modelling creates a bespoke model of radiation exposure based on particular patient data such as imaging scans, body measurements, and medical history. Techniques like computational phantoms based on CT imaging provide highly tailored dose predictions. This method is especially beneficial in paediatric and oncology imaging, where individual factors considerably alter the dose [31].

5. Automated Dose Tracking System

Automated dose-tracking systems are increasingly being employed in clinical settings to track cumulative radiation exposures for patients undergoing repeat imaging. These systems monitor dosages over time, allowing clinicians to review exposure histories and change protocols as needed to lower overall dose. Such systems are particularly advantageous for patients who require frequent imaging, as they can signal instances where cumulative exposure approaches defined safety criteria [32].

6. Artificial Intelligence (AI) driven Tools:

Emerging artificial intelligence (AI) techniques evaluate patient and imaging data to anticipate and optimize radiation dosages for specific patients. AI can help with dose optimization by identifying trends in prior patient data and suggesting lower radiation settings while retaining image quality. This technique, albeit still in development, has promise for real-time dose changes and improving personalised patient care in radiology [33].

Comparative Analysis of Dose Levels in Different Imaging Modalities

Medical imaging provides a visual picture of the body, including organs and tissues, for clinical examination and intervention. Our approach involves sensing and transforming physical signals from the patient's body into medical images. These signals arise from a certain procedure. Radiography transmission through the body includes X-ray projection, fluoroscopy, mammography, and computed tomography. Ultrasound is reflected in the body and utilized for imaging. Precession of a spin system under a strong magnetic field and its use in magnetic resonance imaging [34].

These radiographic imaging methods in this article are x-ray radiography, mammography, fluoroscopy, computed tomography, magnetic resonance imaging, and ultrasound imaging. Medical imaging technology is creating an internal visual representation of the body. This is a non-invasive method of observation without opening the body by surgery [35].

It is used to analyze, diagnose, and treat various medical diseases. Medical imaging technology has developed databases of normal anatomy and physiology, allowing for the detection of abnormalities. Diagnostic imaging technology include both hazards and benefits. This article explores the use of electromagnetic and radio frequency waves in several imaging modalities, including X-ray photography, fluoroscopy, mammography, CT, MRI, and ultrasound. Offer utility and challenge. The views of. This article compares medical imaging modes and technology. The presentation will cover the benefits, hazards, and uses of this technology, as well as specific modes [36].

X- Ray Radiography

Sir Wilhelm Roentgen, a physicist, discovered radiography on November 8, 1895, which revolutionized clinical imaging. Roentgen created the first radiographic image of human anatomy. Radiography, commonly known as radiography, is a specialty of medicine that focuses on analyzing medical pictures. Radiography involves placing an X-ray source on one side of the patient and a flat detector on the opposite side. The X-ray tube produces short-duration X-ray pulses, typically less than 1/2 second. The majority of X-rays contact with the patient, while others pass through and generate an image on the detector. The amount of attenuation caused by scattering and absorption within the patient's tissue affects the uniform distribution of radiographs. The attenuation characteristics of tissues such as bone, soft tissue, and air in the patient's body vary greatly, resulting in an unequal distribution of radiographs sent by the patient. A radiographic image is a result of such X-beam appropriation, and the finder used in radiography might be a visual film (for example, screen film radiography) or an electronic indication framework (computerized radiography). The transmission picture depicts an energy source on one side of the body, which travels through it and is indicated on the other side of the body [37].

Radiography is a technique for transmitting images. Imaging refers to how each point in a picture corresponds to data along a straight line through the patient. Radiography is also an imaging method. Radiographic images can detect a variety of clinical indications, such as fractures, cellular breakdown in the lungs, and cardiovascular infections [38].

Some of the benefits of X -ray Radiography include Non-invasive, quick and painless, support medical and surgical treatment plans, instruct medical personnel to insert catheters or stents into the body to treat tumors or remove blood clots. The risks of X-ray photography include: Ionizing radiation raises the risk of cancer in adulthood, High levels of radiation exposure can cause tissue damage such cataracts, skin redness, and hair loss, medical applications for X-ray photography X-ray photography is utilized for various examinations, including chiropractic and dentistry.

Fluoroscopic X-rays can be used to evaluate blood arteries in the heart and brain, in addition to showing organ movement (e.g. stomach, intestines, colon) [39].

Mammography

Mammography is a type of transmission imaging that uses X-rays to examine the breast. Mammography requires less

X-ray energy than other radiography applications. Therefore, contemporary mammography machines and detection systems are tailored for breast imaging. Mammograms can identify breast cancer in asymptomatic women (screening mammograms) or diagnose breast symptoms (such as lumps (diagnostic mammograms) [40].

Mammography is a specialist X-ray imaging technology that detects breast abnormalities such tumors and calcifications. The mammography above indicates a suspected mass (arrow) that resembles breast malignant tumors, in addition to normal blood arteries and anatomical structures. Special mammography equipment with low X-ray energy, Kedge filter, compression, screen/film detector, anti-scatter grid, and automatic exposure control may provide high-quality, low-dose, and X-ray optimized breast images, as detailed. X-ray mammography is now the preferred approach for screening and early diagnosis of breast cancer due to its high sensitivity, great benefit-risk ratio, and low cost [41].

Some Advantages of mammography includes: Breast cancer detection every 2 years for 20 years, Reduce the risk of dying from breast cancer, among 1000 women who have a mammogram every 2 years for 20 years, 7 cases can be avoided Death, Chemotherapy is needed to reduce the risk of breast cancer, Screening tests can usually detect cancer in the early stages of cancer development. Then it can be treated without chemotherapy, let women understand the health of their breasts, if their mammograms and additional exams do not detect cancer, the vast majority of women (almost 98%) will not develop breast cancer, Screening for breast cancer every 2 years for 20 years may result in [42].



Fig 1: where A is the Mammographs and B is the Mammography's Machine

Fluoroscopy

Fluoroscopy refers to the continuous acquisition of X-beam images throughout time, creating a continuous X-film of the patient. Viewpoint is a projection-based image communication approach similar to continuous radiography3. The fluoroscopy equipment uses an X-beam finder to produce images quickly. Fluoroscopy is used for catheter placement in vessels, identifying different specialists in the GI tract, and other therapeutic uses, such as intrusive treatment methods that require continuous image analysis. Fluoroscopy is used to create X-ray images of physical developments, such as the heart oresophagus [43].

Procedures may include locating a foreign body, doing a blood flow study, injecting anesthetics into the spine or joint, performing orthopedics or urology, inserting a catheter, administering an enema, and implanting a pacemaker. Radiation exposure varies based on patient size and duration of operation. Possible side effects of radiation therapy include skin tissue burns, cancer in old age, and birth abnormalities in pregnant women [44].

Computed Tomography



Fig 2: A-Stenting, B-Implantation, C-Orthopedic surgical view, D-Angiography

Computed tomography started to be utilized in centers in the mid-1970s and was the primary clinical imaging strategy acknowledged by PCs. The CT image is created by pivoting the X-beam tube around the body with the goal that X-beams go through the body at various points. At least one direct identifier clusters confronting the X-beam source to gather transmission projection information. The various information focuses gathered in this manner are combined by a PC into a tomographic image of the patient. The term tomography alludes to a picture (realistic) of a cut (tomo). CT is a transmission innovation that can create image of different squares of tissue in the patient's body. Contrasted and projected pictures, the upside of tomographic image is that they can show the physical construction on the tissue slice (cut) without the upper or lower structure [45].

CT essentially decreased the requirement for exploratory medical procedure, consequently changing clinical practice4. Present day CT scanners can gain 5mm thick tomographic pictures (for example 60 pictures) across a 30cm length of a patient in 10 seconds and uncover the presence of malignant growth, cracked circle, subdural hematoma, aneurysm and a large group of different pathologies [46].

Advantages of computed tomography are: The procedure is simple and painless. The difference lies in the slight difference in physical density. Prevent invasive insertion of arterial catheters and guide wires. Non-invasive and fast. Excellent spatial resolution. And Global vision of veins [47].

Computed tomography is generally used to diagnose trauma, anomalies, and illnesses, Examine several sections of the body, including the brain, sinuses, facial bones, dentistry, spine, cervical spine, hands, wrists, elbows, shoulders, hips, knees, ankles, feet, and kidney tracts [48].

It is generally believed that high-dose ionizing radiation increases the likelihood that a person will continue to develop cancer during their lifetime, but the relationship between low-dose radiation (the order used in standard diagnostic tests) and tumorigenesis does not is clear. (Abdurakhmanovich *et al.,* 2021) [49].

Figure 2.3 shows example of CT where A is Musculoskeletal, B is 3D Fetus, C is Ultrasound guided liver biopsy and D is Color Dopple.



Fig 3: Example of computed Tomography

Comparison between the Medical Imaging Techniques

Medical applications can be compared based on three key elements. The first is picture quality, which includes spatial resolution and contrast. Spatial resolution measures the extent of small objects in an image. Noise refers to how accurately a signal is received. Contrast is the difference between the area of interest in a light or dark image and its background. The second notion is the usability of the system, which can be measured by its cost and real-time availability. The third notion is safety, which includes the impact of ionizing radiation on patients and the effects of warmth on the body (Ababneh *et al*., 2021) [8].

Table 1: Imaging Modalities and properties

Imaging Modalities	Image Quality	System Availability	Safety
X-ray Radiography	1mm special resolution and soft and fluid contrast good	It's don't give real time information and Cost is medium	Heating effect low but have ionizing radiation
Mammography	1mm Special resolution and soft tissue contrast good	It's don't give real time information and Cost is medium	Heating effect low but have ionizing radiation
Fluoroscopy	1 mm special resolution and soft and hard contrast good	It's don't give real time information and Cost is medium	Heating effect low but have ionizing radiation
Computed Tomography	0.5 mm special resolution soft and hard tissue very good contrast, soft tissue contrasts good	It's don't give real time information and Cost is high	Heating effect low but have ionizing radiation.

Conclusion

The operational mechanics and engineering concepts in diagnostic X-ray and CT imaging systems highlight the synergy between physics, engineering, and medical technology. By integrating principles of radiation generation, image acquisition, and system optimization, these modalities provide accurate and reliable diagnostic outcomes. Advancements in detector design, dose management, and computational algorithms continue to enhance safety, image quality, and efficiency. Ultimately, these systems exemplify how engineering innovations drive progress in modern medical imaging and patient care.

References

1. Vassileva J, Rehani MM. Patient dose monitoring and radiation protection in diagnostic imaging: Challenges and solutions. Eur J Radiol. 2020; 126:108921.
2. Ajayi IR, Akinwumiju O. Variations in radiation dose levels during diagnostic imaging in Nigeria. Afri J Med Phy. 2019; 6(3):120-128.
3. Sulieman A, Tamam N, Elnour A, Alkhorayef M, Babikir E, Al-Mohammed HI, *et al*. Effective radiation dose and radiogenic cancer risk during contrast enhanced abdominal computed tomography examinations. Radiation Phy Chem. 2022; 200:110328.
4. Alzimami KS, Ahmed NA, Saleh M. Technical factors affecting radiation dose in computed tomography: A dose optimization approach. Int J Radiol Imag. 2022; 39(4):278-285.
5. Nkubli AJ, Mensah K, Boateng E. Variability in effective doses among X-ray facilities in Ghana: A call for standardization. Radiation Protection and Safety. 2020; 8(1):45-53.
6. Aliyu TM, Musa SA, Umar BA. Assessment of organ equivalent doses in pediatric CT imaging: A Nigerian perspective. J Radiolog Res. 2023; 45(2):98-105.
7. Olowookere OA, Ajibola TS, Oyewole JA. Automated dose monitoring systems in diagnostic radiology: A review of effectiveness in dose optimization. Radiolog Safety J. 2022; 17(2):89-97.
8. Ababneh ZQ, Almusallam M, Alfajri I, Ababneh AM, Aljarrah KM, Tries MA, *et al.* Estimation of the organ and effective dose to patients undergoing medical diagnostic x-ray examinations in saudi arabia. Radiation

Protection Dosimetry. 2021; 194(1):1-8.
9. Hamzian N, Afareydoon S, Ghorbani M, Abrisham SMJ, Roozmand Z, Abdollahi-Dehkordi S, *et al*. Equivalent dose and risk of exposure induced cancer death of different organs due to various image techniques of EOS imaging system. J Biomed Phy Eng. 2021; 11(3):289-296.
10. Al-Hallaq HA, Cerviño L, Gutierrez AN, Havnen-Smith A, Higgins SA, Kügele M, *et al*. AAPM task group report 302: Surface-guided radiotherapy. Medical Physics. 2022; 49(4):e82-e112.
11. Mettler Jr, FA, Mahesh M, Bhargavan-Chatfield M, Chambers CE, Elee JG, Frush DP, *et al*. Patient exposure from radiologic and nuclear medicine procedures in the United States: Procedure volume and effective dose for the period 2006-2016. Radiology. 2020; 295(2):418-427.
12. Fazel M, Soneson E. Current evidence and opportunities in child and adolescent public mental health: A research review. J Child Psychol Psychia. 2023; 64(12):1699-1719.
13. Khramtsov YV, Ulasov AV, Rosenkranz AA, Slastnikova TA, Lupanova TN, Georgiev GP, *et al*. An approach to evaluate the effective cytoplasmic concentration of bioactive agents interacting with a selected intracellular target protein. Pharmaceutics. 2023; 15(2):324.
14. Baptista M, Di Maria S, Vieira S, Pereira J, Pereira M, Vaz P. Organ dose measurements using an adult anthropomorphic phantom and risk estimation of cancer incidence from CBCT exposures. Radiation Phy Chem. 2020; 171:108715.
15. Zewde N, Ria F, Rehani MM. Organ doses and cancer risk assessment in patients exposed to high doses from recurrent CT exams. Eur J Radiol. 2022; 149:110224.
16. Park H, Yoon Y, Kim J, Kim J, Jeong H, Tanaka N, *et al*. Use of clinical exposure index and deviation index based on national diagnostic reference level as dose-optimization tools for general radiography in Korea. Radiation Protection Dosimetry. 2020; 191(4):439-451.
17. Vafapour H. Assessment of the Absorbed Dose Variations in the Thyroid Gland Exposed to Orthopantomography (OPG) while Swallowing: A Novel Approach to Radiation Protection. J Biomed Phy Engineer. 2024; 14(3):267.
18. Vano E, Frija G, Loose R, Paulo G, Efstathopoulos E, Granata C. European Society of Radiology (ESR). Dosimetric quantities and effective dose in medical imaging: A summary for medical doctors. Insights into Imaging. 2021; 12:1-9.
19. Alekseichuk I, Wischnewski M, Opitz A. A minimum effective dose for (transcranial) alternating current stimulation. Brain Stimulation: Basic Transl Clin Res Neuromodulation. 2022; 15(5):1221-1222.
20. Eder H, Schlattl H. Use of effective dose to assess x-ray protective clothing. Journal of Radiological Protection. 2021; 41(4):R140.
21. Alameen S, Tamam N, Awadain S, Sulieman A, Alkhaldi L, Hmed AB. Radiobiological risks in terms of effective dose and organ dose from 18F-FDG whole-body PET/CT procedures. Saudi Journal of Biological Sciences. 2021; 28(10):5947-5951.
22. Niroomand-Rad A, Chiu-Tsao ST, Grams MP, Lewis DF, Soares CG, Van Battum LJ, *et al*. Report of AAPM task group 235 radiochromic film dosimetry: An update to TG-55. Medical Physics. 2020; 47(12):5986-6025.
23. Masjedi H, Zare MH, Keshavarz SN, Razavi-Ratki SK, Alavi F, Shabani M. European trends in radiology: Investigating factors affecting the number of examinations and the effective dose. La Radiologia Medica. 2020; 125:296-305.
24. Tsoulfanidis N, Landsberger S. Measurement and detection of radiation. CRC Press, 2021.
25. Nuszkiewicz J, Woźniak A, Szewczyk-Golec K. Ionizing radiation as a source of oxidative stress-the protective role of melatonin and vitamin D. Int J Mol Sci. 2020; 21(16):5804.
26. Vinnikov VA, Belyakov O. Radiation exposure biomarkers in the practice of medical radiology: Cooperative research and the role of the International Atomic Energy Agency (IAEA) Biodosimetry/Radiobiology Laboratory. Health Physics. 2020; 119(1):83-94.
27. Mahesh M, Ansari AJ, Mettler Jr FA. Patient exposure from radiologic and nuclear medicine procedures in the United States and worldwide: 2009-2018. Radiology. 2022; 307(1):e221263.
28. Mahmoodi M, Chaparian A. Organ doses, effective dose, and cancer risk from coronary CT angiography examinations. Am J Roentgenol. 2020; 214(5):1131-1136.
29. Sterratt D, Graham B, Gillies A, Einevoll G, Willshaw D. Principles of computational modelling in neuroscience. Cambridge University Press, 2023.
30. Sánchez-Nieto B, Medina-Ascanio KN, Rodríguez-Mongua JL, Doerner E, Espinoza I. Study of out-of-field dose in photon radiotherapy: A commercial treatment planning system versus measurements and Monte Carlo simulations. Medical Physics. 2020; 47(9):4616-4625.
31. Bica I, Alaa AM, Lambert C, Van Der Schaar M. From real-world patient data to individualized treatment effects using machine learning: Current and future methods to address underlying challenges. Clin Pharmacol Ther. 2021; 109(1):87-100.
32. Tsalafoutas IA, Kharita MH, Al-Naemi H, Kalra MK. Radiation dose monitoring in computed tomography: Status, options and limitations. Physica Medica. 2020; 79:1-15.
33. Huynh E, Hosny A, Guthier C, Bitterman DS, Petit SF, Haas-Kogan DA, *et al*. Artificial intelligence in radiation oncology. Nature Reviews Clinical Oncology. 2020; 17(12):771-781.
34. Kattar N, Migneron M, Debakey MS, Haidari M, Pou AM, McCoul ED. Advanced computed tomographic localization techniques for primary hyperparathyroidism: A systematic review and meta-analysis. JAMA Otolaryngology-Head and Neck Surgery. 2022; 148(5):448-456.
35. Aggarwal R, Sounderajah V, Martin G, Ting DS, Karthikesalingam A, King D, *et al*. Diagnostic accuracy of deep learning in medical imaging: A systematic review and meta-analysis. NPJ Digital Medicine. 2021; 4(1):65.
36. Zieleskiewicz L, Markarian T, Lopez A, Taguet C, Mohammedi N, Boucekine M, Azurea Network. Comparative study of lung ultrasound and chest computed tomography scan in the assessment of severity of confirmed COVID-19 pneumonia. Intensive Care

Medicine. 2020; 46:1707-1713.
37. Wang Y, Jia S, Wei M, Peng L, Wu Y, Liu X. Research progress on solidification structure of alloys by synchrotron X-ray radiography: A review. Journal of Magnesium and Alloys. 2020; 8(2)::396-413.
38. Pandiyan V, Masinelli G, Claire N, Le-Quang T, Hamidi-Nasab M, De Formanoir C, *et al*. Deep learning-based monitoring of laser powder bed fusion process on variable time-scales using heterogeneous sensing and operando X-ray radiography guidance. Additive Manufacturing. 2022; 58:103007.
39. Ho CS, Chen YP, Fan TY, Kuo CF, Yen TY, Liu YC, *et al*. Application of deep learning neural network in predicting bone mineral density from plain X-ray radiography. Archives of Osteoporosis. 2021; 16:1-12.
40. Jochelson MS, Lobbes MB. Contrast-enhanced mammography: State of the art. Radiology. 2021; 299(1):36-48.
41. Sogani J, Mango VL, Keating D, Sung JS, Jochelson MS. Contrast-enhanced mammography: Past, present, and future. Clinical Imaging. 2021; 69:269-279.
42. Yala A, Mikhael PG, Strand F, Lin G, Smith K, Wan YL, *et al*. Toward robust mammography-based models for breast cancer risk. Science Transl Med. 2021; 13(578):eaba4373.
43. Shalom NE, Gong GX, Auster M. Fluoroscopy: An essential diagnostic modality in the age of high-resolution cross-sectional imaging. World J Radiol. 2020; 12(10):213.
44. Debreceni D, Janosi K, Bocz B, Turcsan M, Lukacs R, Simor T, *et al*. Zero fluoroscopy catheter ablation for atrial fibrillation: A systematic review and meta-analysis. Front Cardiovas Med. 2023; 10:1178783.
45. Sera T. Computed tomography. Transparency in Biology: Making the Invisible Visible, 2021, l167-1187.
46. Withers PJ, Bouman C, Carmignato S, Cnudde V, Grimaldi D, Hagen CK, *et al*. X-ray computed tomography. Nature Reviews Methods Primers. 2021; 1(1):18.
47. Pontone G, Scafuri S, Mancini ME, Agalbato C, Guglielmo M, Baggiano A, *et al*. Role of computed tomography in COVID-19. Journal of Cardiovascular Computed Tomography. 2021; 15(1):27-36.
48. Hsieh J, Flohr T. Computed tomography recent history and future perspectives. Journal of Medical Imaging. 2021; 8(5):052109-052109.
49. Abdurakhmanovich KO, Servetovna AA, Zairovna AU. Multispiral computed tomography for prostate diseases. Central Asian Journal of Medical and Natural Science. 2021; 2(2):9-11.